



CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

Memo No.STM/STORE/QUOTE/192/2022-23 QUOTATION NOTICE Date:28.02.2023

Sealed QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Microbiology (Virology Unit) **School of Tropical Medicine, Kolkata.**

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 01.03.2023

The last date of submission of QUOTATION is **09.03.2023 up to 12.00 pm** and the same will be opened on the same date at **2.00 pm** in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder Company
GST Registration certificate
IT Return of the last financial year

The financial quotation Should be quoted as per the following table only :-

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Quotation Drop box STM Main Building (2nd floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.

No hand written document accepted only printed documents submitted.

SL No	Item
1	Maxiamp 0.1ml. Low Profile Tube Strips with Cap Ultra Clear flat Cat number-611030 125 per pack

Rang 28/2/2023

Medical Superintendent cum Vice Principal,
School of Tropical Medicine, Kolkata

MSVP
School of Tropical Medicine
Kolkata



CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

Memo No.STM/STORE/QUOTE/193/2022-23 QUOTATION NOTICE Date:28.02.2023

Sealed QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Microbiology (Virology Unit) School of Tropical Medicine, Kolkata. The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 01.03.2023

The last date of submission of QUOTATION is 09.03.2023 up to 12.00 pm and the same will be opened on the same date at 2.00 pm in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder Company
GST Registration certificate
IT Return of the last financial year

The financial quotation Should be quoted as per the following table only :-

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Quotation Drop box STM Main Building (2nd floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.

No hand written document accepted only printed documents submitted.

SL No	Item
1	Herpes simplex virus and Varicella Zoster virus Real Time PCR Detection Kit (1 pack/96 reactions)

Specifications for RTPCR kit for Herpes virus and Varicella Zoster virus

1. Should be in a kit format,
2. Should be CE_IVD certified with documentary evidence
3. Bids must include external quality assurance results for performance evaluation
4. The kit should include a heterologous Internal Control which can be used for both sample preparation control and as a PCR inhibition control to confirm the integrity of the reagents of the kit
5. Should have positive control in the kit and not to be purchased separately
6. Kits must be compatible with ABI 7500Fast Dx, CFX96, which documentary evidence must be provided from the manufacturer along with programming guidebooks / manuals from the manufacturer, without needing any additional software to be purchased.
7. The kits should have extended stability of at least 12 months from the date of manufacture and documentary evidence from the manufacturer
8. Should be validated from CSF, Skin lesions ,Blood samples by the manufacturer
9. Kits should be based on real-time PCR technology, for the simultaneous detection and differentiation of herpes simplex virus 1 (HSV-1), herpes simplex virus 2 (HSV-2), varicella-zoster virus (VZV) DNA
10. The kit should contain reagents in multiple aliquots to avoid multiple freeze thaw cycles
11. Must have a lower limit of detection not exceeding 20 DNA copies per reaction
12. Should use common channels FAM, ROX, Cy5, HEX, JOE or VIC for detection and not specific machine dependent channels for highest sensitivity
13. The kit should not cross-react with other similar symptom causing viruses.

Ran 28/2/23

Medical Superintendent cum Vice Principal,
School of Tropical Medicine, Kolkata

School of Tropical Medicine
Kolkata



CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

KolkataMemo No.STM/STORE/QUOTE/194/2022-23 QUOTATION NOTICE Date:28.02.2023

Sealed QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Microbiology (Virology Unit) School of Tropical Medicine, Kolkata.
The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 01.03.2023

The last date of submission of QUOTATION is 09.03.2023 up to 12.00 pm and the same will be opened on the same date at 2.00 pm in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder Company
GST Registration certificate
IT Return of the last financial year

The financial quotation Should be quoted as per the following table only :-

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Quotation Drop box STM Main Building (2nd floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.

No hand written document accepted only printed documents submitted.

SL No	Item
1	Cyto megallo (CMV) virus Real Time PCR Detection kit (1 pack/96 reactions)

Specifications for RTPCR kit for Cyto megallo (CMV) virus

1. Should be in a kit format,
2. Should be CE_IVD certified with documentary evidence
3. Bids must include external quality assurance results for performance evaluation
4. The kit should include a heterologous Internal Control which can be used for both sample preparation control and as a PCR inhibition control to confirm the integrity of the reagents of the kit
5. Should have positive control in the kit and not to be purchased separately
6. Kits must be compatible with ABI 7500Fast Dx, CFX96, which documentary evidence must be provided from the manufacturer along with programming guidebooks / manuals from the manufacturer, without needing any additional software to be purchased.
7. The kits should have extended stability of at least 12 months from the date of manufacture and documentary evidence from the manufacturer
8. Should be validated from CSF, Skin lesions ,Blood samples by the manufacturer
9. Kits should be based on real-time PCR technology, for the simultaneous detection and differentiation of herpes simplex virus 1 (HSV-1), herpes simplex virus 2 (HSV-2), varicella-zoster virus (VZV) DNA
10. The kit should contain reagents in multiple aliquots to avoid multiple freeze thaw cycles
11. Must have a lower limit of detection not exceeding 20 DNA copies per reaction
12. Should use common channels FAM, ROX, Cy5, HEX, JOE or VIC for detection and not specific machine dependent channels for highest sensitivity
13. The kit should not cross-react with other similar symptom causing viruses.

Handwritten signature and date: 28/2/2023

Medical Superintendent cum Vice Principal,
School of Tropical Medicine, Kolkata

School of Tropical Medicine
Kolkata