



# CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

Memo No. STM/STORE/QUOTE/102/2025-26 QUOTATION NOTICE

Date:12.09.2025

Sealed Short-QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Department of Cord Blood Bank, **School of Tropical Medicine, Kolkata.**

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 13.09.2025

The last date of submission of QUOTATION is **18.09.2025 up to 12.00 pm** and the same will be opened on the same date in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment  
PAN card of the Bidder /Bidder Company  
GST Registration certificate  
IT Return of the last financial year


**The financial quotation Should be quoted as per the following table only :-**

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

- Quotation Drop box STM Main Building (2<sup>nd</sup> floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.
- This Quotation validate up to three (03) months.
- No hand written document accepted only printed documents submitted.


Sl. No	Items	Quantity
1.	Cord Blood Processing Kit (Specification attached)	100 units

  
Medical Superintendent cum Vice Principal,  
School of Tropical Medicine, Kolkata

MSVP  
School of Tropical Medicine  
Kolkata

**Technical specification of Cord Blood Processing kit  
(CBP Kit) for the Cord Blood Bank, STM, Kolkata:**

1. The CBP Kit must consist of three bags integrally connected by tubing, two of which are PVC bags – one larger bag & one smaller bag and a third smallest bag which is the Cryo bag /freezing bag.
2. The larger bag should have a fill volume of 250ml with an attached spike for transferring content from a different collection bag. It must have one leak proof access port for injecting chemicals inside or taking out sample.
3. Other smaller bag should have a fill volume of 150ml. It must have one leak proof access port for injecting chemicals inside or taking samples from the bag.
4. The third integrally attached bag (Cryo bag/Freezing bag) consists of EVA material or approved material for storing stem cells at ultra low temperature (minus 196°C) of liquid nitrogen. It must have two compartments of 20 ml & 5 ml capacity with inter-communication.
5. Height of this Cryo bag/Freezing bag should be 76 mm and width should be 90 mm. The smallest bag should have 2 (two) ports (for connecting spike for transplant) of the bag one connected with the 20 ml part and other connected with the 5 ml part). This Cryo bag/ Freezing bag along with at least 210 mm integral tube segment should fit inside the metal cassette with dimension 96-97 mm (Length) X 90-91 mm (Width) X 9-10 mm (thickness) for preservation.
6. Temperature range of storage of the CBP Kit before processing should be between + 10°C to + 40°C.
7. Easily manipulative clamp must be present on every connecting tubing from bag to bag and from bag to spike and bag to entry or exit ports.
8. The kit should have an expiry of at least 2 years from the date of manufacture.
9. The kit manufacturing process should follow the national and or international standards like ISO 13485 , so that the bags are pyrogen free, sterilised, animal pathogenicity test passed, can withstand the shearing force during centrifugation etc.
10. Specimen of the CBP kit must be provided to the tender committee for precise verification of the size, shape, measurement and other specifications.
11. Other requirement (if approved and order is placed for purchase):-Each supply lot/batch has to be provided with Certificate of Analysis (quality assurance certificate) otherwise the kit will not be accepted.
12. Approval letter of manufacturing this product must be provided.

  
08/09/2025





# CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

Memo No. STM/STORE/QUOTE/103/2025-26 QUOTATION NOTICE

Date:12.09.2025

Sealed Short-QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Department of Cord Blood Bank, **School of Tropical Medicine, Kolkata.**

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 13.09.2025

The last date of submission of QUOTATION is **18.09.2025 up to 12.00 pm** and the same will be opened on the same date in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment  
PAN card of the Bidder /Bidder Company  
GST Registration certificate  
IT Return of the last financial year


**The financial quotation Should be quoted as per the following table only :-**

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

- Quotation Drop box STM Main Building (2<sup>nd</sup> floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.
- This Quotation validate up to three (03) months.
- No hand written document accepted only printed documents submitted.

Sl. No	Items	Quantity
1.	Urinalysis Strip - 2 Parameter Strip (Microalbumin Creatinine Urine Strips)	05pack

  
Medical Superintendent cum Vice Principal,  
School of Tropical Medicine, Kolkata

MSVP  
School of Tropical Medicine  
Kolkata