GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073 Phone: 033 2212 3695/96/97 Fax: 033 2212 3698

Website: stmkolkata.org

Memo No. STM/STORE/QUOTE/191

QUOTATION NOTICE

Dt. 17.02.2020

Sealed QUOTATION are invited from the Reputed Company/Direct Importers/Authorized Distributor firms for supply of items as mentioned below at the Department of **CCU School of Tropical Medicine**, Kolkata.

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 18.02.2020

The last date of submission of QUOTATION is **26.02.2020 up to 1 pm** and the same will be opened on the same date at 2 pm in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents:

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder company
GST Registration certificate
IT Return of any two of last three financial years
Valid P.Tax Certificate

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Payment Fund: RSBY Fund.

1. Air Purifier

Air Purifier Technical Specification:

- Minimum 7 Stage Filteration system
- Build with Lung Filter, H1N1 Filter, HEPA Filter, Activated Carbon Filter, Anti-Bacterial filter and Ionizer.
- Intelligent Air Quality Monitor.
- Noise Level Area: 66dBA or Less.
- Applicable Area: More than 2210 Sq Ft.
- Clean Air Delivery Rate (CADR): 780m³/h or More.
- Anion Concentration: 10 million/cm³ or More.
- Control Operation by Touch Panel & Remote.
- Inter Lock & Child lock Protection.
- Minimum 1 year warranty.

Medical Superintendent cum Vice Principal School of Tropical Medicine, Kolkata

Medical Superintendent-Cum-Vice-Principal
Carmichael Hospital for Tropical Diseases
School of Tropical Medicine

Kolkata

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073 Phone: 033 2212 3695/96/97 Fax: 033 2212 3698

Website: stmkolkata.org

Memo No. STM/STORE/QUOTE/192

QUOTATION NOTICE

Dt. 17.02.2020

Sealed QUOTATION are invited from the Reputed Company/Direct Importers/Authorized Distributor firms for supply of items as mentioned below at the Department of **Cord Blood Bank School of Tropical Medicine**, Kolkata.

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 18.02.2020

The last date of submission of QUOTATION is **26.02.2020 up to 1 pm** and the same will be opened on the same date at 2 pm in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents:

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder company
GST Registration certificate
IT Return of any two of last three financial years
Valid P.Tax Certificate

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products , technical brochures along with Test certificate produced and physical sampling thereof.

Payment Fund: State Budget Fund

1. Cord Blood Processing Bag.

Encl: Technical Specification attached.

Medical Superintendent cum Vice Principal School of Tropical Medicine, Kolkata

2020

Medical Superintendent-Cum-Vice-Principal
Parmichael Hospital for Propical Diseases
School of Tropical Medicine

Kolkata

Technical specification of "Cord Blood Processing Bag for the Cord Blood Bank, STM, Kolkata:

- 1. The CBP Kit must consist of three bags integrally connected by tubing, two of which are PVC bags one larger bag & one smaller bag and a third smallest bag which is the Cryo bag /freezing bag.
- 2. The larger bag should have a fill volume of 200-250 ml with an attached spike for transferring content from a different collection bag. It must have one leak proof access port for injecting chemicals inside or taking out sample. One additional entry port to this bag is preferred for injecting chemicals only through the nozzle of the syringe without using needle with very narrow calibre connecting tubing for administration of Cryoprotectant(s).
- 3. Other smaller bag should have a fill volume of 150-175 ml. It must have one leak proof access port for injecting chemicals inside or taking samples from the bag.
- 4. The third integrally attached bag (Cryo bag/Freezing bag) consists of EVA material or approved material for storing stem cells at ultra low temperature (minus 196°C) of liquid nitrogen. It should be a 3 dimensional or 2-dimensional with a dual compartment 25ml bag consist of 20ml & 5ml capacity.
- 5. Length of this Cryo bag/Freezing bag is 85-90 mm & Height 60-70 mm (However the total height is 70-75 mm including the height of the two ports of the bag one connected with the 20ml part and other connected with the 5ml part). This Cryo bag/Freezing bag along with 100-210 mm integral tube segment should fit inside the metal cassette with dimension 91-92mm (Length) X 84-85mm (Width) X 8-9mm (thickness) for preservation.
- 6. Temperature range of storage should be minus 196 °C to plus 40° C.
- 7. Easily manipulative clamp must be present on every connecting tubing from bag to bag and from bag to spike and bag to entry or exit ports.
- 8. The kit should have an expiry of at least 2 years from the date of manufacture.
- 9. The kit manufacturing process should follow the national and or international standards like ISO etc. so that the bags are pyrogen free, sterilised, animal pathogenicity test passed, can withstand the shearing force during centrifugation etc.
- 10. Should have US FDA/ Indian FDA approval certificate.
- 11. Other requirement (if approved and order is placed for purchase):-Each supply lot/batch has to be provided with Certificate of Analysis (quality assurance certificate) otherwise the kit will not be accepted.



GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073 Phone: 033 2212 3695/96/97 Fax: 033 2212 3698

Website: stmkolkata.org

Memo No. STM/STORE/QUOTE/199

QUOTATION NOTICE

Dt. 17.02.2020

Sealed QUOTATION are invited from the Reputed Company/Direct Importers/Authorized Distributor firms for supply of items as mentioned below at the Department of **Lab Medicine School of Tropical Medicine**, Kolkata.

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 18.02.2020

The last date of submission of QUOTATION is **26.02.2020 up to 1 pm** and the same will be opened on the same date at 2 pm in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents:

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder company
GST Registration certificate
IT Return of any two of last three financial years
Valid P.Tax Certificate

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Payment Fund: State Budget Fund

1. Microtips (1000ul)

Medical Superintendent cum Vice Principal School of Tropical Medicine, Kolkata

Medical Separate and ent-Cum-Vice-Principal Carmichael Hospital for Tropical Diseases
School of Tropical Medicine
Kolkata



GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073 Phone: 033 2212 3695/96/97 Fax: 033 2212 3698

Website: stmkolkata.org

Memo No. STM/STORE/QUOTE/200

QUOTATION NOTICE

Dt. 17.02.2020

Sealed QUOTATION are invited from the Reputed Company/Direct Importers/Authorized Distributor firms for supply of items as mentioned below at the Department of **Microbiology School of Tropical Medicine**, Kolkata.

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 18.02.2020

The last date of submission of QUOTATION is **26.02.2020 up to 1 pm** and the same will be opened on the same date at 2 pm in presence of interested bidder , if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

Valid Trade License / Enlistment
PAN card of the Bidder /Bidder company
GST Registration certificate
IT Return of any two of last three financial years
Valid P.Tax Certificate

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Payment Fund: State Budget Fund

1. Vortex machine. (Tarson or REMI)

17.02.2020

Medical Superintendent cum Vice Principal School of Tropical Medicine, Kolkata

Medical Superintendent-Cum-Vice-Principal Carmichael Hospital for Tropical Diseases School of Tropical Medicine Kolkata